

ANALIZA BIAŁEK



100012241

## Laboratory of Analysis

**Client:** RIVER PEPTIDES  
**Order number:** 100012241  
**Name:** SS-31 SS101121  
**Number of samples:** 1 vial

**Batch number:** -----  
**Recived date:** 2026-01-27  
**Start date:** 2026-02-06  
**End date:** 2026-02-10

## Sample identification

Substance	Method of taking a sample	Form
SS-31	Sample provided by customer	White powder in glass vial

## Results

Substance	Sample name	Content	Purity
SS-31	SS-31 SS101121	10,02mg	99%

## Research comment

The test sample was applied in the amount of 3  $\mu$ l to the BioBasic 8 Thermo Scientific column. During the course of the analysis with the method a linear gradient in the solutions of H<sub>2</sub>O and Acetonitrile with the addition of 0.1% trifluoroacetic acid, a peak of the sought substance appeared. Its content, when compared to the reference chromatogram, is 10,02mg. Height the measurement error is + - 0,05mg. The actual state of the sample sent before the test is presented on the photo below

## Photo of the test sample




### Password

A93LNF74


**Analiza Białek sp. z o.o.**  
**KRS:** 0000966168  
**NIP:** 8992922970  
**REGON:** 521751875

**Headquarters and Laboratory:**  
ul. Duńska 9, 54-427 Wrocław  
Budynek DELTA

**Lab technician:**  
Analysis and research specialist  
mgr Arkadiusz Zajęc



## Chromatogram



## Method

Time	H2O + 0,1% TFA	ACN + 0,1% TFA
0min	60%	40%
2min	60%	40%
12min	5%	95%
15min	5%	95%

## Password

A93LNF74